

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	250205181		2025.05.02	2027.05.02
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.33	0.28	1.54	23.45-24.19	
1 L - QC	18.11	0.45	2.46	22.96-24.12	
5 ml - QC	19.15	0.32	1.66	24.48-25.31	
QC Analysis by: Carlos A. Ulloa		Reviewed by: Nicole Thibodeau			
Date: 2025-05-20		Date: 2025-05-20			
Signature: Carlos A. Ulloa		Signature: Nicole Th			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.