

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	260306181		2026.06.03	2028.06.03
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.50	0.20	1.08	23.79-24.31	
1 L - QC	18.15	0.71	3.92	22.66-24.51	
5 ml - QC	18.43	0.26	1.42	23.61-24.29	
QC Analysis by: Carlos A. Villoz		Reviewed by: MIKE O'HARA			
Date: 2026-07-24		Date: 2026/07/30			
Signature: Carlos A. Villoz		Signature: Mike O'Hara			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.