

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	260306351		2026.06.03	2028.06.03
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	35.67	0.72	2.01	45.44-47.31	
1 L - QC	34.99	1.34	3.84	43.73-47.23	
5 ml - QC	35.93	1.19	3.30	45.16-48.25	
QC Analysis by: Carlos A. Ulloa		Reviewed by: MIKE O'HARA			
Date: 2026/07/24		Date: 2026/07/30			
Signature: Carlos A. Ulloa		Signature: MIKE O'HARA			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.