

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	260308031		2026.08.03	2028.08.03
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	3.28	0.19	5.88	4.02-4.52	
1 L - QC	3.18	0.23	7.18	3.84-4.43	
5 ml - QC	3.30	0.08	2.38	4.19-4.40	
QC Analysis by: Carlos A. Ulloa		Reviewed by: MIKE O'HARA			
Date: 2026/08/25		Date: 2026-08-25			
Signature: Carlos A. Ulloa		Signature: MIKE O'HARA			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.