

Mouse Embryo Assay Certificate of Analysis

Product:	Expert Holding Micropipettes
----------	------------------------------

Catalog #: HP.130B.30

Lot Number: HHP241225

Expiration Date: 2027-12-25

Sterilization:

The products were gamma irradiated at a minimum dose of 27.2 kGy and a maximum dose of 24.8 kGy and 12/20/2024. The validated sterile assurance let

maximum dose of 34.8kGy on 12/29/2024. The validated sterile assurance level is 10⁻⁶.

Assays	Specification	Results
Endotoxin Assay (LAL Test) ¹	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) ²	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	100%

¹Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

All assays were performed by an independent laboratory.

Approved by:

Date: 1/9/2025

²For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.