

Mouse Embryo Assay Certificate of Analysis

Product: Expert Hold	ing Micropipettes
----------------------	-------------------

 Catalog #:
 HP.180B.30

 Lot Number:
 HHP250225

 Expiration Date:
 2028-02-25

Sterilization:

The products were gamma irradiated at a minimum dose of 27.3 kGy and a maximum dose of 35.0kGy on 2/28/2025. The validated sterile assurance level is 10^{-6} .

Assays	Specification	Results
Endotoxin Assay (LAL Test) ¹	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) ²	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	95%

¹Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

All assays were performed by an independent laboratory.

Approved by:

Date: 3/13/2025

²For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.