

# Mouse Embryo Assay Certificate of Analysis

**Product:** Intelligent ICSI Micropipettes

**Catalog #:** IC.3.50.30.A

**Lot Number:** HIC250225

**Expiration Date:** 2028-02-25

**Sterilization:** The products were gamma irradiated at a minimum dose of 27.3 kGy and a maximum dose of 35.0kGy on 2/28/2025. The validated sterile assurance level is  $10^{-6}$ .

Assays	Specification	Results
Endotoxin Assay (LAL Test) <sup>1</sup>	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) <sup>2</sup>	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	95%

<sup>1</sup>Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

<sup>2</sup>For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.

All assays were performed by an independent laboratory.

Approved by:



Date: 3/13/2025