Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	250708181		2025.08.07	2027.08.08
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	18.39	0.31	1.68	23.50-24.30	
1 L - QC	18.54	0.16	0.88	23.89-24.31	
5 ml - QC	18.86	0.22	1.18	24.23-24.81	
QC Analysis by: Gaby U. Ulloa		Reviewed by: Bryan Smith Lowe			
Date: 2025/09/22		Date: 2025/09/25			
Signature: Gaby U. Ulloa		Signature: Bryan Smith Lowe			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.