Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	250708351		2025.08.07	2027.08.08
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	36.44	0.57	1.56	46.63-48.10	
1 L - QC	35.34	1.32	3.73	44.23-47.66	
5 ml - QC	36.56	0.50	1.37	46.87-48.18	
QC Analysis by: Cdos A. Ulloa		Reviewed by: Bryan Saint-Louis			
Date: 2025/09/26		Date: 2025/09/29			
Signature: Cdos A. Ulloa		Signature: Bryan Saint-Louis			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.