Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	251109031		2025.09.11	2027.09.11
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	3.27	0.15	4.68	4.05-4.45	
1 L - QC	3.29	0.03	0.97	4.23-4.32	
5 ml - QC	3.30	0.21	6.26	4.02-4.55	

QC Analysis by: Gilda A. Ulloa	Reviewed by: Bryan Scott Lewis
Date: 2025/09/19	Date: 2025/09/25
Signature: Gilda A. Ulloa	Signature: Bryan Scott Lewis

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.