Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	251205351		2025.05.12	2027.05.12

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	35.69	0.27	0.77	46.04-46.75
1 L - QC	36.58	1.26	3.46	45.90-49.19
5 ml - QC	37.06	0.59	1.59	47.41-48.94

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Thibodeau
Date: 2025-05-20	Date: 2025.05.20
Signature: Carlos A. Ulloa	Signature: Nicole Th

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.