

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	252006181		2025.06.20	2027.06.20
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.67	0.21	1.12	24.00-24.55	
1 L - QC	18.39	0.18	0.97	23.68-24.14	
5 ml - QC	18.66	0.28	1.48	23.89-24.61	
QC Analysis by: Carlos A. W/102		Reviewed by: Bryan Saint-Louis			
Date: 2025-07-25		Date: 2025/07/25			
Signature: Carlos A. W/102		Signature: Bryan Saint-Louis			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.