

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	252306351		2025.06.23	2027.06.23
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	35.32	0.57	1.60	45.18-46.65	
1 L - QC	35.62	0.68	1.91	45.41-47.19	
5 ml - QC	35.99	0.51	1.41	46.12-47.44	
QC Analysis by: Carlos A. Ullor		Reviewed by: Bryan Saint-Louis			
Date: 2025-07-25		Date: 2025/07/25			
Signature: Carlos A. Ullor		Signature: Bryan Saint-Louis			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.