Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	241206031		2024.06.12	2026.06.12
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	3.23	0.08	2.43	4.10-4.30	
1 L - QC	3.11	0.11	3.57	3.89-4.18	
5 ml - QC	3.12	0.15	4.71	3.86-4.24	
QC Analysis by: Carlos A. Ulloa		Reviewed by: Nicole Thibodeau			
Date: 2024-11-27		Date: 2024-12-02			
Signature: Carlos A. Ulloa		Signature: Nicole Th			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.