

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	261304181		2026.04.13	2028.04.13
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.15	0.39	2.14	23.09-24.10	
1 L - QC	18.14	0.26	1.44	23.25-23.92	
5 ml - QC	17.97	0.51	2.84	22.69-24.02	
QC Analysis by: <i>Colin A. Upton</i>		Reviewed by: <i>MIKE O'HARA</i>			
Date: <i>2026/05/29</i>		Date: <i>2026-06-01</i>			
Signature: <i>Colin A. Upton</i>		Signature: <i>MIKE O'HARA</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.