

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	261304351		2026.04.13	2028.04.13
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	36.62	0.31	0.84	47.21-48.01	
1 L - QC	35.76	0.87	2.42	45.36-47.61	
5 ml - QC	36.72	0.71	1.93	46.81-48.66	
QC Analysis by: <i>Carlos A. Ulloa</i>		Reviewed by: <i>MIKE O'HARA</i>			
Date: <i>2026/05/29</i>		Date: <i>2026-06-01</i>			
Signature: <i>Carlos A. Ulloa</i>		Signature: <i>MIKE O'HARA</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.