

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	260209181		2026.09.02	2028.09.02
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.47	0.04	0.22	23.96-24.07	
1 L - QC	18.17	0.64	3.50	22.79-24.44	
5 ml - QC	17.41	0.23	1.33	22.33-22.93	
QC Analysis by: <i>Carlos A. 11/10/22</i>		Reviewed by: <i>Mike O'Hara</i>			
Date: <i>2026/09/29</i>		Date: <i>2026/10/07</i>			
Signature: <i>Carlos A. 11/10/22</i>		Signature: <i>Mike O'Hara</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.

CONFIDENTIAL AND PROPRIETARY — Property of Nexpring Health, HTI. This document contains confidential information. Unauthorized use, disclosure, copying, or distribution is prohibited. E