

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	260309351		2026.09.03	2028.09.03
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	34.94	0.56	1.59	44.70-46.15	
1 L - QC	33.79	1.09	3.23	42.51-45.35	
5 ml - QC	33.62	0.96	2.84	42.46-44.95	
QC Analysis by: Carlos A. Ulloa		Reviewed by: Mike O'Hara			
Date: 2026/09/29		Date: 2026/10/07			
Signature: Carlos A. Ulloa		Signature: Mike O'Hara			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.

CONFIDENTIAL AND PROPRIETARY — Property of Nexpring Health, HTI. This document contains confidential information. Unauthorized use, disclosure, copying, or distribution is prohibited.