Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	240403351		2024.03.04	2026.03.05
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	34.93	0.34	0.97	44.97-45.85	
1 L - QC	35.54	1.09	3.07	44.78-47.61	
5 ml - QC	34.95	1.18	3.39	43.89-46.97	
QC Analysis by: Gaby A. Ulloa		Reviewed by: Nicole Thibodeau			
Date: 3-21-2024		Date: 2024.03.22			
Signature: Gaby A. Ulloa		Signature: Nicole Th			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.