

Mouse Embryo Assay Certificate of Analysis

Product: Brilliant Biopsy Micropipettes

Catalog #: BP.25BLU.30

Lot Number: HBP250825

Expiration Date: 2028-08-25

Sterilization: The products were gamma irradiated at a minimum dose of 25.3 kGy and a maximum dose of 36.3kGy on 8/30/2025. The validated sterile assurance level is 10^{-6} .

Assays	Specification	Results
Endotoxin Assay (LAL Test) ¹	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) ²	% expanded to blastocysts	
Control	80%	95%
This Lot	80%	100%

¹Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

²For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.

All assays were performed by an independent laboratory.

Approved by:



Date: 9/10/2025