

Mouse Embryo Assay  
Certificate of Analysis

**Product:** Expert Holding Micropipettes

**Catalog #:** HP.100N.30

**Lot Number:** HHP250825

**Expiration Date:** 2028-08-25

**Sterilization:** The products were gamma irradiated at a minimum dose of 25.3 kGy and a maximum dose of 36.3kGy on 8/30/2025. The validated sterile assurance level is  $10^{-6}$ .

| Assays                                  | Specification             | Results           |
|-----------------------------------------|---------------------------|-------------------|
| Endotoxin Assay (LAL Test) <sup>1</sup> | ≤0.5 EU/device            | <0.03125EU/device |
| Mouse Embryo Assay (MEA) <sup>2</sup>   | % expanded to blastocysts |                   |
| Control                                 | 80%                       | 95%               |
| This Lot                                | 80%                       | 100%              |

<sup>1</sup>Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

<sup>2</sup>For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.

All assays were performed by an independent laboratory.

Approved by:



Date: 9/10/2025