

Mouse Embryo Assay
Certificate of Analysis

<u>Product:</u>	Expert Holding Micropipettes
<u>Catalog #:</u>	HP.120N.30
<u>Lot Number:</u>	HHP251125D
<u>Expiration Date:</u>	2028-11-25
<u>Sterilization:</u>	The products were gamma irradiated at a minimum dose of 27.7 kGy and a maximum dose of 34.8kGy on 11/30/2025. The validated sterile assurance level is 10^{-6} .

Assays	Specification	Results
Endotoxin Assay (LAL Test) ¹	≤ 0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) ²	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	100%

¹Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

²For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.

All assays were performed by an independent laboratory.

Approved by:



Date: 12/8/2025