

## Mouse Embryo Assay Certificate of Analysis

Product:	Expert Holding Micropipettes	
Catalog #:	HP.150B.30	
Lot Number:	HHP250525	

Expiration Date: 2028-05-25

Sterilization: The products were gamma irradiated at a minimum dose of 28.3 kGy and a maximum dose of 36.9kGy on 5/31/2025. The validated sterile assurance level is  $10^{-6}$ .

Assays	Specification	Results
Endotoxin Assay (LAL Test) <sup>1</sup>	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) <sup>2</sup>	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	100%

<sup>&</sup>lt;sup>1</sup>Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

All assays were performed by an independent laboratory.

Approved by:

Date: <u>6/23/2025</u>

QF-71-01-05 Rev. A

<sup>&</sup>lt;sup>2</sup>For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.