

Mouse Embryo Assay
Certificate of Analysis

Product: Intelligent ICSI Micropipettes

Catalog #: IC.1.50.35.A

Lot Number: HIC251025D

Expiration Date: 2028-10-25

Sterilization: The products were gamma irradiated at a minimum dose of 25.8 kGy and a maximum dose of 35.6kGy on 10/24/2025. The validated sterile assurance level is 10^{-6} .

Assays	Specification	Results
Endotoxin Assay (LAL Test) ¹	≤0.5 EU/device	<0.03125EU/device
Mouse Embryo Assay (MEA) ²	% expanded to blastocysts	
Control	80%	100%
This Lot	80%	100%

¹Utilizes the LAL gel clot assay with a sensitivity of 0.03125EU/mL

²For a valid assay, at least 80% of 2-cell stage control embryos should develop to expanded blastocyst within 72 hours. The product is considered non-toxic if 80% of the embryos exposed to it develop to expanded and/or hatching blastocysts within 72 hours.

All assays were performed by an independent laboratory.

Approved by:



Date: 11/6/2025