

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2105
Production date	02 July 2021
Expiry date	31 January 2022
Total volume (mL)	1580

	Results	Specifications
pH	7,31	7,20-7,60
Osmolality	301	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	84	≥ 80 %

Release date	19 July 2021
--------------	--------------

Released by	Vanessa Bierlich
-------------	------------------

Signature

P.P. 