

**Certificate of Analysis
GM501 SpermMobil**

Batchnumber	SMOBIL-2106
Production date	11 August 2021
Expiry date	28 February 2022
Total volume (mL)	1590

	Results	Specifications
pH	7,26	7,20-7,60
Osmolality	298	280-320 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,50 EU/mL
Sperm survival Test 4 hours exposure	93	≥ 80 %

Release date	27 August 2021
--------------	----------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann