


Certificate of Analysis GM501 SpermMobil

Batchnumber	SMOBIL-2307
Production date	05 September 2023
Expiry date	31 March 2024
Total volume (ml)	2900

Parameter	Results	Specifications
pH	7,32	7,20-7,60
Osmolality	300	280-320 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
Sterility test (according to the current Ph.Eur. 2.6.1 / USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to the test medium)	96	≥ 80 %

Release date	20 September 2023
Released by	Nicole Geiger
Signature	

REV 08_00 / 2023-09-15