XYCLONE® RED-i RESEARCH LASER (V5.12)

100 Cummings Center, Suite 465E, Beverly MA 01915 978.921.2050, 88.323.0503, Fax: 978.921.0250 www.hamiltonthorne.com, info@hamiltonthorne.com

The XYClone® includes laser module with 40x or 20x objective, 4 turret adapters, video camera, c-mount adapter, proprietary laser drilling software, desktop or laptop computer and accessories. Microscope is not included.

Dimensions	H	\mathbf{W}	D	
	in. (mm)	in. (mm)	in. (mm)	lb. (kg)
XYClone:	1.65 (42)	1.1 (28)	2.1 (54)	0.5 (0.22)
Controller:	3.8 (95)	2.5 (63.5)	1.8 (46)	1.0 (0.45)
MiniTower:	11.5 (292)	3.7 (92.6)	11.4 (290)	11.57 (5.26))
Monitor: (with stand)	20.2 (513.5)	21.9 (556)	7.1 (180.3)	10.1 (4.6)
Laptop:	0.9 (23.45)	14.8 (376.9)	10.1 (255.2)	4.71 (2.14)

Electrical	Desktop	Monitor	Laptop
Input Voltage:	100/240 VAC	110-240 VAC	100-240 VAC
Power:	300 W	72 W (max)	90W
Line Frequency:	50/60 Hz	50/60 Hz	50/60 Hz

Laser

Firing:

Type: 1460 nm, Infrared Solid State Diode
Maximum Power: @ Focus = 300 mW (Class I)
Laser Modes: Single, Double, Staccato
Laser produces single laser shots

Adjustable Laser Power (%) and Pulse Length (µs)

Fire laser by mouse or footswitch

Double: Provides access to two independent Single Laser panels

Each panel has its own power and pulse setting and laser fire

button

Fire laser by mouse or footswitch Staccato: Optional: Pulse length: 1 - 3000 μs

Power: 1 - 100%

Repetition rate per sec: 1 - 1000 Mean power maximum: 90 mW By mouse (Optional: Foot switch firing

Target Marker: Circle or arrow, adjustable "Blink Time" after firing.

Isotherm Rings, showing peak temps and hole size. Select

which rings to display.

Crosshairs: Activation, size and color set by user (used for positioning

embryo)

Laser Alignment: Aligned and locked at factory. No on-site physical laser

alignment required

Target: Adjust target alignment on-screen

RED-i Target Locator

Source: Red LED built into laser module

Alignment: X-Y mechanical adjustment to center of laser beam using

Isotherm Rings as guide

Intensity: On-screen intensity adjustment

Objective

Scale Bar:

Standard: 40x 0.50 N.A. or 20x 0.40 N.A., I.R. (High transmission in both the visible and near-infrared, long working distance)

Scale Calibration: Performed interactively on-screen

Calibrate and save multiple objectives Scale bar graphic overlay automatically adjusts based on

calibrated objective. May be saved with images/video.

Video

Camera: Standard: High resolution digital color

Optional: Analog black & white Digital: 1360 x 1024 pixels

Image Area: Digital: 1360 x 1024 pixels
Optional NTSC: 640 x 480 pixels

Zoom: 2x, 4x, and 8x with user defined image panning

Illumination: Microscope, image on screen

Laser Safety

EN IEC 60825-1:2014: Class 1 21 CFR 1040:10: Class I

Image Capture and Storage Utility

Capture and store unlimited images. Images stored in user-selected JPG, BMP, or TIF format. Capture unlimited thumbnail images and select which to save. Automatic image capture on laser fire. Manual or automatic file naming. Images may be saved with graphic overlay.

HAMILTON THORNE

Image Annotation Tools

Unlimited automatic image labels may be stored and enabled. Freehand text, circles, rectangles, lines, and image measures may be added to captured images.

Video Capture

Capture and store real time and time lapse video in .avi format. Ability to set maximum recording length. Manual and automatic naming options. Open and play saved videos within program.

Measurement Toolbox

Tools allow measurement of various embryo parameters on captured images. Each measure visible as graphic overlay, including length in micron *Zona*: 5 zona thickness measurements; Mean & Standard Deviation

Embryo: 2 diameter measurements; Mean & Standard Deviation; Blastomere

count

Pronuclei: 2 diameter measurements for two separate pronuclei; Mean for each

pronuclei

Drill: 5 hole size measurements; Mean & Standard Deviation Ruler: 5 user-defined measurements; Mean & Standard Deviation

Reports and Output

Data Input: Data from keyboard

Data from measurements ASCII Import critical fields

Report: Ova/embryo information, procedure/protocol information, choice

of 2 images plus evaluation data or 4 images

Output: Printout of report

Report stored in JPG format

ASCII output of all numerical and alphanumeric fields in TXT &

MER formats

Computer (subject to change)

 Type:
 Dell MiniTower / SFF
 Dell Laptop

 OS:
 Windows 10 Pro (64-bit)
 Windows 10 Pro (64-bit)

 CPU:
 Intel Core i7
 Intel Core i7

 Memory:
 8 GB
 8 GB DDR3

 Drives:
 1 TB GB HD
 500 GB HD

8x DVD+/-RW SATA External 8x DVD +/-RW Display: 24" Flat Panel Widescreen (16:10) 15.6" HD Anti-glare

Ports: 6 USB 3.0 (2 on front) 3 USB 3.0

4 USB 2.0 (2 on front) (one with PowerShare)

 1 Serial
 1 HDMI

 2 Display Port
 1 Display Port

 2 PS/2
 SD Memory card reader

1 Line-in (stereo/microphone)
1 Line-out (headphone/speaker)

Network: 10/100/1000 Ethernet

Combo Stereo headphone & Microphone jack
10/100/1000 Ethernet

1 RJ45 port (used by camera) Wireless LAN

Mouse: Wired Wired Keyboard: Wired Integrated

Specifications subject to change without notice.

Doc. # SS-1223 Rev. F 16-Mar.-2021