

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	250511181		2025.11.05	2027.11.06
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.10	0.11	0.61	23.39-23.67	
1 L - QC	17.75	0.19	1.10	22.82-23.32	
5 ml - QC	18.30	0.39	2.15	23.28-24.30	
QC Analysis by: Carlos A. Ullas		Reviewed by: PAUL BRANNER			
Date: 2026/01/30		Date: 2026/01/30			
Signature: Carlos A. Ullas		Signature: Paul Branner			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.