

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	250611351		2025.11.06	2027.11.06
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	35.56	0.26	0.74	45.88-46.57	
1 L - QC	35.49	1.06	2.98	44.76-47.50	
5 ml - QC	34.28	0.84	2.44	43.48-45.65	
QC Analysis by: Carlos A. Ullora		Reviewed by: PAUL BARNER			
Date: 2026/01/30		Date: 2026/01/30			
Signature: Carlos A. Ullora		Signature: Paul Barner			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.