

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	260902351		2026.02.09	2028.02.09
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	35.07	0.39	1.10	45.09-46.09	
1 L - QC	34.81	1.13	3.23	43.78-46.71	
5 ml - QC	37.58	0.09	0.25	48.73-48.97	
QC Analysis by: <i>Bobby G. Wilson</i>		Reviewed by: <i>Bryan Saint-Louis</i>			
Date: <i>2026/03/18</i>		Date: <i>2026/03/18</i>			
Signature: <i>Bobby G. Wilson</i>		Signature: <i>Bryan Saint-Louis</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.