

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	261002031		2026.02.10	2028.02.10
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	3.17	0.11	3.42	3.98-4.27	
1 L - QC	3.09	0.25	8.18	3.69-4.34	
5 ml - QC	3.44	0.21	6.07	4.19-4.74	
QC Analysis by: <i>[Signature]</i>		Reviewed by: <i>MIKE O'HARA</i>			
Date: <i>2026/04/22</i>		Date: <i>2026-04-23</i>			
Signature: <i>[Signature]</i>		Signature: <i>[Signature]</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.