

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	261802181		2026.02.18	2028.02.18
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	18.13	0.12	0.68	23.41-23.73	
1 L - QC	18.29	0.19	1.02	23.54-24.02	
5 ml - QC	18.21	0.67	3.68	22.80-24.54	
QC Analysis by: <i>Andy U. W. [Signature]</i>		Reviewed by: <i>Bayan Saint-Louis [Signature]</i>			
Date: <i>2026/03/18</i>		Date: <i>2026/03/18</i>			
Signature: <i>Andy U. W. [Signature]</i>		Signature: <i>Bayan Saint-Louis [Signature]</i>			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.