Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads+®	Part Number	710259

Type	Range	Lot #	Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	172808031	2021-02-01	2023-02-01

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1L-Bio	2.95	0.22	7.51	3.54 to 4.12
1L-QC	2.96	0.11	3.77	3.71 to 4.00
5ml-QC	3.00	0.05	1.56	3.84 to 3.96

Bio Analysis by:	D.H. Doug Hamilton	QC Analysis by:	David Lynch
Date:	2021.04.26	Date:	2021-03-22
Signature:		Signature:	

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Technical Note. Please contact Hamilton Thorne for more information.