Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

*** AMENDED ***

CERTIFICATE OF ANALYSIS

Product	accu-beads+®	Part Number	710257

Type	Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	192711181	2021/01/13	02/2023

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1L-Bio	18.52	0.43	2.30	23.53 to 24.63
1L-QC	18.09	0.34	1.88	23.07 to 23.96
5ml-QC	17.75	0.32	1.82	22.66 to 23.50

Bio Analysis by: Tom Kenny, Sr.	QC Analysis by: David Lynch
Date: 2021/02/25	Date: 2021-02-25
Signature:	Signature:

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Technical Note. Please contact Hamilton Thorne for more information.