Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	220212351		2022.12.02	2024.12.02
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	36.17	1.24	3.42	45.41-48.63	
1 L - QC	35.07	1.19	3.38	44.04-47.13	
5 ml - QC	34.88	1.08	3.09	43.94-46.74	

QC Analysis by: Gaby U. Ulloa	Reviewed by: NICOLE THIBODEAU
Date: 04-05-2023	Date: 2023.04.06
Signature: Gaby U. Ulloa	Signature: Nicole T

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.