Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	220707181		2022.07.07	2024.11.18
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	18.43	0.21	1.11	23.69-24.22	
1 L - QC	18.32	0.41	2.23	23.28-24.35	
5 ml - QC	17.82	0.64	3.59	22.33-23.99	

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Thibodeau
Date: 2023-02-27	Date: 2023-02-27
Signature: Carlos A. Ulloa	Signature: Nicole Th

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.