Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

# CERTIFICATE OF ANALYSIS

| Product | accu-beads®+ | Part Number | 710258 |
|---|---|---|---|

| Concentration Type | Acceptable Range | Lot # | | Date of Manufacture | Expiration Date |
|---|---|---|---|---|---|
| 35 M/ml | ± 5.0 M/ml | 220802351 | | 2022.02.08 | 2024.04.29 |
| **Test Type and Quantity** | **20 µm Fixed Coverslip Chamber Calibration** | **S.D.** | **CV%** | **Calculated Hemacytometer Calibration Range*** | |
| 1 L - Bio | 36.80 | 0.81 | 2.20 | 46.79-48.90 | |
| 1 L - QC | 35.03 | 1.55 | 4.43 | 43.52-47.56 | |
| 5 ml - QC | 34.81 | 0.71 | 2.03 | 44.33-46.17 | |
| QC Analysis by: Carlos A. Ulloa | | Reviewed by: Douglas Hamilton | | | |
| Date: 2022-08-18 | | Date: 2022.08.22 | | | |
| Signature: Carlos A. Ulloa | | Signature: Douglas Hamilton | | | |

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.