Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	220905181		2022.05.09	2024.09.06
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	18.24	0.36	1.98	23.25-24.19	
1 L - QC	18.19	0.83	4.59	22.56-24.73	
5 ml - QC	18.50	0.32	1.76	23.62-24.47	
QC Analysis by: Carlos A. Ulloa		Reviewed by: Nicole Thibodeau			
Date: 2023-03-16		Date: 2023.03.16			
Signature: Carlos A. Ulloa		Signature: Nicole Th			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.