Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	221403181	2022.03.14	2024.05.10

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.15	0.08	0.44	23.49-23.70
1 L - QC	18.66	0.54	2.90	23.56-24.96
5 ml - QC	17.64	0.30	1.68	22.55-23.32

QC Analysis by: Carlos A. Ulloa	Reviewed by:
Date: 2022-09-06	Date: 2022.09.07
Signature:	Signature:

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.