Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	221403351		2022.03.14	2024.09.06
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	34.95	0.07	0.20	45.35-45.53	
1 L - QC	34.75	1.36	3.90	43.41-46.93	
5 ml - QC	35.62	0.35	0.98	45.84-46.75	

QC Analysis by: Carla A. Allen	Reviewed by: DHH
Date: 2022-11-16	Date: 2022-11-16
Signature: Carla A. Allen	Signature: DHH

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.