Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	223011181	2022.11.03	2024.12.02

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.61	0.40	2.13	23.68-24.71
1 L - QC	18.27	0.50	2.76	23.10-24.41
5 ml - QC	18.05	0.60	3.32	22.68-24.24

QC Analysis by: Carlos H. Ulloa	Reviewed by: Nicole Thibodeau
Date: 04-05-2023	Date: 2023-04-06
Signature: Carlos H. Ulloa	Signature: Nicole Th

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.