Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	230604181	2023.04.06	2025.04.07

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.50	0.35	1.90	23.59-24.50
1 L - QC	18.59	0.36	1.94	23.69-24.16
5 ml - QC	18.90	0.48	2.52	23.95-25.19

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Thibodeau
Date: 2023-06-30	Date: 2023.06.30
Signature: Carlos A. Ulloa	Signature: Nicole T

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.