Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	230604351	2023.04.06	2025.04.07

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	34.88	0.36	1.04	44.87-45.81
1 L - QC	34.44	0.30	0.86	44.38-45.15
5 ml - QC	35.02	0.68	1.94	44.65-46.41

QC Analysis by: Carlos N. Ulloa	Reviewed by: Nicole Thibodeau
Date: 2023-06-30	Date: 2023.06.30
Signature: Carlos N. Ulloa	Signature: Nicole Th

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.