Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	230709031		2023.09.07	2025.09.07
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	3.30	0.03	1.06	4.25-4.34	
1 L - QC	3.50	0.03	0.85	4.51-4.59	
5 ml - QC	3.38	0.20	6.0	4.13-4.65	

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Thibodeau
Date: 2023-09-15	Date: 2023.09.19
Signature: Carlos A. Ulloa	Signature: Nicole T

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.