Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	231209351	2023.09.12	2025.09.28

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	35.67	0.70	1.95	45.47-47.28
1 L - QC	33.93	0.22	0.65	43.28-44.39
5 ml - QC	35.38	0.45	1.26	45.42-46.58

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Thibodeau
Date: 01-09-2024	Date: 2024.01.09
Signature: Carlos A. Ulloa	Signature: Nicole Thibodeau

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.