Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	232806181	2023.08.06	2025.07.11

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.50	0.38	2.04	23.56-24.54
1 L - QC	18.05	0.39	2.19	22.95-23.97
5 ml - QC	18.57	0.27	1.47	23.78-24.49

QC Analysis by: Dean PoP	Reviewed by: Nicole Thibodeau
Date: 08/29/2023	Date: 2023.08.30
Signature:	Signature:

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.