Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	232806351		2023.06.28	2025.07.11
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	35.00	0.17	0.49	45.28-45.72	
1 L - QC	33.47	0.63	1.89	42.69-44.33	
5 ml - QC	34.70	0.66	1.89	44.26-45.96	
QC Analysis by: Dean Pop		Reviewed by: Nicole Thibodeau			
Date: 08/29/2023		Date: 2023.08.30			
Signature: Dean Pop		Signature: Nicole T			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.