Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	232908181	2023.08.29	2025.09.07

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.32	0.30	1.64	23.42-24.21
1 L - QC	18.36	0.30	1.66	23.47-24.26
5 ml - QC	18.49	0.47	2.53	23.42-24.64

QC Analysis by: [signature]	Reviewed by: Nicole Thibodeau
Date: 2023-10-20	Date: 2023.10.20
Signature: [signature]	Signature: [signature]

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.