Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	232908351		2023.08.29	2025.09.07
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	35.39	0.26	0.74	45.67-46.35	
1 L - QC	36.07	0.66	1.82	46.03-47.74	
5 ml - QC	34.59	1.67	4.83	42.79-47.13	
QC Analysis by: Emily A. Wilson		Reviewed by: Nicole Thibodeau			
Date: 2023-10-20		Date: 2023.10.20			
Signature: Emily A. Wilson		Signature: Nicole Th			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.